
# **MORPHOLINE**



#### 1 PERFORMANCE

1) Measuring range : 2-22ppm Number of pump strokes 1(100mL)

2) Sampling time : 1 minute/1 pump stroke

3) Detectable limit : -

4) Shelf life : 3 years 5) Operating temperature :  $15 \sim 25$  °C

6) Reading : Graduations printed on the tube are calibrated by Ammonia

at 1 pump stroke and Morpholine concentration is

determined by using a conversion chart at 1 pump stroke.

7) Colour change : Pale purple  $\rightarrow$  Pale Yellow

## 2. RELATIVE STANDARD DEVIATION

RSD-low: 10% RSD-mid.: 5% RSD-high: 5%

#### 3 CHEMICAL REACTION

By reacting with Phosphoric acid, PH indicator is discoloured.  $C_4H_9\,NO+H_3\,PO_4\to (R_3\,NH)_3\,PO_4$ 

# 4. CALIBRATION OF THE TUBE

PERMEATION TUBE METHOD

### INTERFERENCE AND CROSS SENSITIVITY

Substance	Interference	Coexistence
Amines	Similar stain is produced.	Higher readings are given.

