

Tube No.
110S

GASOLINE



1. PERFORMANCE


- 1) Measuring range : 0.05-0.6 %
Number of pump strokes 1(100mℓ)
(This detector tube is calibrated by Hexane.)
- 2) Sampling time : 1 minute/1 pump stroke
3) Detectable limit : 50 ppm
4) Shelf life : 3 years
5) Operating temperature : 10 ~ 40 °C
6) Temperature compensation : Necessary (See "TEMPERATURE CORRECTION TABLE")
7) Reading : Direct reading from the scale calibrated by 1 pump stroke
8) Colour change : Orange → Dark green

2. RELATIVE STANDARD DEVIATION

RSD-low : 10 % RSD-mid. : 10 % RSD-high : 5 %

3. CHEMICAL REACTION

Chromium oxide is reduced.



4. CALIBRATION OF THE TUBE

GAS CHROMATOGRAPHY

5. INTERFERENCE AND CROSS SENSITIVITY

Substance	Interference	ppm	Coexistence
Benzene	FIG.1	Similar stain is produced.	400 Higher readings are given.
Toluene	FIG.2	〃	800 〃
Xylene		〃	2,000 〃
Acetylene		Whole layer is discoloured to Dark green.	〃
Ethylene		Similar stain is produced.	〃
Cyclohexane		〃	〃
Paraffin hydrocarbons		〃	〃

In presence of Alcohols, Ketones or Esters less than 6 %, the accuracy of readings is not affected.




FIG.1 Influence of Benzene




FIG.2 Influence of Toluene

TEMPERATURE CORRECTION TABLE

Scale Readings (%)	True Concentration (%)				
	10°C (50°F)	15°C (59°F)	20°C (68°F)	30°C (86°F)	40°C (104°F)
0.6	0.73	0.65	0.60	0.52	0.48
0.5	0.60	0.55	0.50	0.45	0.42
0.4	0.48	0.44	0.40	0.37	0.35
0.3	0.35	0.32	0.30	0.28	0.27
0.2	0.22	0.21	0.20	0.19	0.18
0.1	0.10	0.10	0.10	0.10	0.10