
HYDROGEN SULPHIDE



1. PERFORMANCE

1) Measuring range 2-60 ppm 1-30 ppm Number of pump strokes $1/2(50 \text{m} \ell)$ $1(100 \text{m} \ell)$ 2) Sampling time $1/2(50 \text{m} \ell)$ $1(100 \text{m} \ell)$ 3) Detectable limit $1/2(50 \text{m} \ell)$ $1/2(50 \text{m} \ell)$ 1/2(5

4) Shelf life : 3 years 5) Operating temperature : $0 \sim 40^{\circ}$ C

6) Reading : Direct reading from the scale calibrated by 1 pump stroke

7) Colour change : White→Pale brown

2. RELATIVE STANDARD DEVIATION

RSD-low: 10% RSD-mid.: 5% RSD-high: 5%

3. CHEMICAL REACTION

By reacting with Lead acetate (II), Lead sulphide is produced. H₂S + Pb (CH₃COO)₂ → PbS + 2CH₃COOH

4. CALIBRATION OF THE TUBE

PERMEATION TUBE METHOD

5. INTERFERENCE AND CROSS SENSITIVITY

Substance		Interference	ppm	Coexistence
Sulphur dioxide	FIG.1	The accuracy of readings is not affected.	10	Higher readings are given.
Mercaptans	FIG.2	"	300	"
Nitrogen dioxide	FIG.3	"	2	Lower readings are given.

(NOTE)

In case of 1/2 pump strokes, following formula is available for the actual concentration.

Actual concentration = $2 \times$ Reading value

