DIMETHYL ETHER



1. PERFORMANCE

1) Measuring range 0.01-1.2%Number of pump strokes $1(100\text{m}\ell)$

2) Sampling time : 1.5 minutes/1 pump stroke

3) Detectable limit \vdots 10 ppm 4) Shelf life \vdots 3 years 5) Operating temperature \vdots 0 \sim 40 $^{\circ}$ C

6) Reading : Direct reading from the scale calibrated by 1 pump stroke

7) Colour change : Orange → Dark brown

2. RELATIVE STANDARD DEVIATION

RSD-low: 10% RSD-mid.: 10% RSD-high: 10%

3. CHEMICAL REACTION

Potassium dichromate is reduced. $CH_3OCH_3 + Cr^{6+} + H_2SO_4 \rightarrow Cr^{3+}$

4. CALIBRATION OF THE TUBE

GAS CHROMATOGRAPHY

5. INTERFERENCE AND CROSS SENSITIVITY

Substance	Interference	ppm	Coexistence
Acetylene		3%	Whole reagent is changed to Brown.
Propane		0.2%	"
Other organic gases or vapours except Halogenated hydrocarbons	Similar stain is produced.		Higher readings are given.