# Tube No. 600SP

# CARBON MONOXIDE



### 1. PERFORMANCE

1) Measuring range : 5-100 ppm2) Detectable limit : 3 ppm3) Shelf life : 2 years4) Operating temperature  $: 0 \sim 40 \, \text{°C}$ 

5) Reading : Direct reading from the scale printed on the tube

6) Colour change : Yellow→Dark brown

#### 2. RELATIVE STANDARD DEVIATION

RSD-low: 10% RSD-mid.: 10% RSD-high: 10%

#### 3. CHEMICAL REACTION

Pottasium disulphate palladate (II) is reduced and Palladium is liberated.

 $CO + K_2Pd(SO_3)_2 \rightarrow Pd$ 

## 4. CALIBRATION OF THE TUBE

STANDARD GAS CYLINDER METHOD

#### 5. PRESET/MEASUREMENT CONDITIONS

#### PRESET CONDITION

1) 2nd pressure : 0.6 kg/cm<sup>2</sup> (60 Kpa)

2) Flow rate  $\therefore$  2.0  $\ell$ /min.

#### MEASUREMENT CONDITION

Sampling time : 2.0 minutes